



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/662,697	09/15/2003	William J. Boyle	ACS 65470 (2309D)	9777
24201	7590	09/24/2004	EXAMINER	
FULWIDER PATTON LEE & UTECHT, LLP HOWARD HUGHES CENTER 6060 CENTER DRIVE TENTH FLOOR LOS ANGELES, CA 90045			WEBB, SARAH K	
			ART UNIT	PAPER NUMBER
			3731	
DATE MAILED: 09/24/2004				

Please find below and/or attached an Office communication concerning this application or proceeding.

# Office Action Summary

Application No.

10/662,697

Applicant(s)

BOYLE ET AL.

Examiner

Sarah K Webb

Art Unit

3731

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

## Status

- 1) ☒ Responsive to communication(s) filed on 15 September 2003.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

## Disposition of Claims

- 4) ☒ Claim(s) 35-62 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 35-62 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

## Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

## Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

## Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)  
Paper No(s)/Mail Date \_\_\_\_\_
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date. \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: \_\_\_\_\_

## DETAILED ACTION

### ***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

1. Claims 35-40,42-50,52-54,58-62 are rejected under 35 U.S.C. 102(e) as being anticipated by US Patent No. 6,171,327 to Daniel et al.

Daniels illustrates a catheter system in Figures 20-23 that is designed for recovery of an embolic filter (21) that is disposed on a guide wire (26). The retrieval device includes an inner catheter (172 in Figure 20 or 372 in Figure 23) that extends distally beyond a recovery sheath (151). The inner catheter can be constructed to have either greater or less column strength than the recovery sheath, as evidenced by the various configurations of Figures 20 and 23. Each catheter has a control handle attached to its proximal end, and the handles are illustrated in Figures 24-26. Control handle 702 is connected to the proximal end of the recovery sheath (151) and control handle 710 is connected to the proximal end of the inner catheter (372).

Inner catheter (372) can be locked onto the guide wire (26) by way of a threaded connection between the handle (710) and a locking mechanism that includes a guide wire clamp (720) and a collet (718). The recovery sheath control handle (702) is locked with the inner catheter control handle (710) by a stop (708) that prevents the handles

Art Unit: 3731

(702,710) from becoming separated but allows the handles to slide relative to one another.

Regarding claims 36,46,56, and 57, the language "may be up to", "may be up to approximately", and "may extend up to" is significantly broad to include any length less than the stated dimension. Therefore, the Daniel device meets this limitation, since the recovery sheath is clearly shorter than the inner catheter.

Daniel discloses steps of using the device in column 10 that include advancing the inner catheter and recovery sheath over a guide wire, locking the inner catheter to the guide wire, advancing the recovery sheath over the filter to collapse it, and removing the entire system from the patient's body.

### ***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

2. Claims 41 and 51 are rejected under 35 U.S.C. 103(a) as being unpatentable over Daniel in view of US Patent No. 5,201,757 to Heyn et al.

Daniel includes all the limitations of claims 41 and 51, except that the position of the handles is switched so that control handle of the recovery sheath is coaxially disposed within the lumen of the control handle of the inner catheter. Heyn discloses a device with control handles for sheaths that move relative to one another. Heyn teaches that the control handle (60) for the inner catheter (44) can be disposed within

Art Unit: 3731

the control handle (56) of the outer sheath (20). It would have been obvious to one of ordinary skill in the art at the time the invention was made to simply rearrange the control handles of Daniel so that the control handle of the inner catheter is disposed within the lumen of the recovery sheath handle, as Heyn teaches that this is an alternate way to configure control handles of relatively moving sheaths.

### ***Conclusion***

3. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. US Patent 6,569,184 to Huter is significantly similar to the claimed invention. US Patent 6,371,971 to Tsugita discloses various embodiments of a filter retrieval device.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Sarah K Webb whose telephone number is (703) 605-1176. The examiner can normally be reached on Mon-Fri 8-4:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Anhthuan T. Nguyen can be reached on (703) 308-2154. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.


Art Unit: 3731

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

SKW

09/21/2004

SKW

  
DAVID O. REIP  
PRIMARY EXAMINER